## **Digital Sample Freezing Platform - DSFP**






Digital/Virtual Freezing of Samples

Design of platform accessible to everyone (Go-back-in-time capabilities)





Small scale testing of digital freezing

## **Digital Sample Freezing Methodology**




#### Interface for processing recorded LC-HRMS data at www.norman-data.eu







### **Compound Selection Module**










Compounds with prior-knowledge on spectral behavior:

- ✓ Ionization pre-selection
  - ✓ Adduct pre-selection
- Qualifier fragments in evaluation of chromatograms module

### Sample Selection Module



## Sample Selection Module Meta-Data retrieval

241 ⊖

■ UoA LC-ESI-QTOF Wastewater effluent Athens Greece 07.03.2015 TREMEPOL

Greece

Water-Waste water-Municipal and Industrial

2015-03-07

University of Athens

**TREMEPOL** 

Latitude (Decimal): 37.94051

Longitude (Decimal): 23.5883

Precision of coordinates: Precise (range 1-10 m) Laboratory Name: Laboratory of Analytical Chemistry

Organization Country: Greece

Laboratory City: Athens

Contact Person: Nikolaos Thomaidis Type of data source: Monitoring data Type of monitoring: Surveillance

Instrument manufacturer: Bruker maxis Impact

Analytical Column: Thermo Acclaim RSLC C18 2.2um, 2.1x100mm

Reconstitution solvents: Methanol/Water 50/50

Injection volume (ul): 5 Column temperature (C): 25

Mobile phases A 90:10 water:methanol with 0.01% formic acid and 5mM ammonium formate; B methanol with 0.01% formic acid and 5mM ammonium formate

Gradient: 99% A at 0-1 min; 61% A at 3 min; 0.1% A at 14-16 min; 99% A at 16.1-20 min


Flow (mL/min): 200 uL/min at 0-3 min; 400 uL/min at 14 min; 480 uL/min at 16-19 min; 200 uL/min at 19.1-20 min

MS Scan Range: 50-1000

Declared mass accuracy MS1: 2 Declared mass accuracy MS2: 3


Fragmentation method: CID

#### Results




Selection of results for deeper inspection from experts


### Module for Evaluation of Chromatograms



## 1-Click Interactive Spatial Distribution Map




#### **Batch Mode Module**




Easy 4-step batch-mode screening


### Heatmap Module




Blue color indicates that compound was detected in the corresponding sample White color indicates that compound was not detected


- DEET the most common active ingredient in insect repellents,
- MS/MS spectra at three different collision energies








- Task: the presence of DEET during JDS3 survey
- Tools: predicted RTI (±), mass accuracy (±), specific fragments
- JDS-58-Arges unambiguously positive for DEET



 JDS\_3 - results: - 53 out of 68 sampling points were positive for DEET



 EMBLAS - results: - 32 out of 55 sampling points were positive for DEET

